

# CERTIFICATE OF ANALYSIS



## Petrifilm® Yeast and Mold Count Plates

Item No. 700002121, 700002128 or 700006809

Catalog No. 6407, 6417 or 6445

**Lot No.** 346H67

**Manufacture Date:** 2025-08-26

**Expiration Date:** 2027-02-24

**This CofA supersedes the prior CofA dated Sept 25, 2025**

Organism Tested	Minimum Growth	Result
<i>Aspergillus niger</i> , M6	Productivity Ratio $\geq$ 0.7	Pass
<i>Paecilomyces</i> species, M10	Productivity Ratio $\geq$ 0.7	Pass
<i>Botrytis</i> species, M97	Productivity Ratio $\geq$ 0.7	Pass
<i>Hansenula anomala</i> , Y28	Productivity Ratio $\geq$ 0.7	Pass
<i>Saccharomyces cerevisiae</i> , Y41	Productivity Ratio $\geq$ 0.7	Pass
<i>Enterococcus faecalis</i>	No Growth	Pass
<i>Acinetobacter</i> species	No Growth	Pass

### Chemical and Physical Testing

Parameter	Specification	Result	Pass
pH	7.0 $\pm$ 1.0	6.78	Yes

### Contamination Check

Test Plan	Result
96 Randomized plates per batch tested per a statistical sampling plan	Pass

This material complies with the Neogen specifications for this product construction, and applicable criteria for routine quality control and microbiological performance. Neogen is certified to ISO 9001 through an independent agency.

Manufactured for Neogen Corporation by Solventum, 601 22nd Ave. South, Brookings, SD 57006

*Gabrielle Jackson*

Gabrielle Jackson  
Date: 2025-10-28

CofA-700002121,  
700002128 or  
700006809  
FS-COA-0228  
Rev: 3



620 Lesher Place • Lansing, MI 48912 • 800.234.5333 (U.S. and Canada) • 517.372.9200  
The Dairy School • Auchincruive • Ayr, KA6 5HU • Scotland, UK • +44 (0) 1292 525 600