

CERTIFICATE OF ANALYSIS



Campylobacter Enrichment Broth, 250 grams

Item No. 700002285
Catalog No. CE250

Lot No. 2507AY01

Manufacture Date: 2025-07-21

Expiration Date: 2027-01-19

Chemical and Physical Testing

Parameter	Specification	Result	Pass
Appearance Color	Tan	Tan	Yes
pH	7.0 ± 0.2	7.14	Yes

Microorganism Testing*

Organism Tested	Growth Specification	Pass
<i>Campylobacter jejuni</i> ATCC 33291	>1e ⁵ CFU/mL	Yes
<i>Campylobacter jejuni</i> ATCC 29428	>1e ⁵ CFU/mL	Yes
<i>Campylobacter coli</i> ATCC 43478	>1e ⁵ CFU/mL	Yes
<i>Campylobacter lari</i> BAA-1060	>1e ⁵ CFU/mL	Yes
<i>Escherichia coli</i> ATCC 25922	No Growth	Yes
<i>Escherichia coli</i> ATCC 8739	No Growth	Yes
<i>Proteus mirabilis</i> ATCC 29906	No Growth	Yes

*The following results were obtained from cultures incubated in CE250 at 41.5°C for 24hs.

Terminal Irradiation Processing Parameters

This product has been subjected to Gamma irradiation as a terminal process. The minimum and maximum delivered met established specifications for this product.

This letter certifies that the product described above has been tested and found to comply with Neogen's product performance specifications applicable to subject device at time of manufacture. Visit the Neogen Food Safety website, www.Neogen.com/foodsafety for additional information. Neogen is certified to ISO 9001 through an independent agency.

Manufactured for Neogen Corporation by Alpha Biosciences, 3651 Clipper Mill Rd. Baltimore, MD 21211.

Gabrielle Jackson

Gabrielle Jackson
Date: 2025-10-13



CofA-700002285

FS-COA-0262

Rev: 2

620 Lesher Place • Lansing, MI 48912 • 800.234.5333 (U.S. and Canada) • 517.372.9200

The Dairy School • Auchincruive • Ayr, KA6 5HU • Scotland, UK • +44 (0) 1292 525 600